



**INSTRUMENTS**

# **35-3010RKA-03 Sample-Draw Detector Head**

*Part Number: 71-0134RK*

*Released: 11/2/06*

*Revision: P1*

---

## Specifications

Table 1 lists specifications for the 35-3010RKA-03.


**Table 1: Specifications**

|                                 |   |
|---------------------------------|---|
| Target Gases & Detection Ranges | Combustible Gas, 0 - 100% LEL<br>Oxygen, 0-25% volume<br>Carbon Monoxide, 0 - 300 ppm |
| Input Power                     | 24 VDC  |
| Current Draw                    | 275 mA  |
| Output Signals                  | 4-20 mA each channel  |
| Construction (housing)          | Fiberglass/polyester (NEMA 4X)  |
| Dimensions                      | 15.44 in. H x 12.55 in. W x 8.31 in. D  |
| Weight                          | 14 lbs.   |
| Sampling Method                 | Sample-draw   |
| Sample Flow                     | 1.2 SCFH (nominal)  |
| Response Time                   | 90% in 30 seconds   |
| Accuracy                        | ±5% of detection range  |
| Repeatability                   | ±2% of detection range  |

---

## Description

This section describes the components of the 35-3010RKA-03 sample-draw detector head. The sample-draw detector consists of the housing, flow system, and detection system.



**Figure 1: Sample-draw Detector Head Component Location**

### Housing


The sample-draw detector's fiberglass housing is weather- and corrosion-resistant. It is suitable for installation where general purpose equipment is in use. The housing door is hinged on the left side and is secured by two latches on the right side.

Four mounting feet are attached to the back of the housing (one at each corner). Use the mounting feet to install the housing to a vertical surface. Use the two conduit hubs on the bottom of the housing to make wiring connections.

An aluminum subpanel is mounted to the interior of the housing. The sample-draw detector's internal components are mounted to the subpanel.

## Flow System

The sample-draw detector's flow system consists of the INLET fitting, hydrophobic filter, charcoal filter, pump, flowmeter, bypass valve, status lights, pressure switch, flow block, and EXHAUST fitting (see Figure 1). Figure 2 illustrates how the gas sample moves through the flow system.



**Figure 2: Sample-Draw Detector Flow Diagram**

### *INLET fitting*

The INLET fitting on the bottom of the housing allows the gas sample to enter the sample-draw detector. The INLET fitting accepts 1/4 in. rigid tubing. See the Installation section on page 11 to connect tubing to the INLET fitting.

### *Hydrophobic Filter*

The hydrophobic filter is to the left of the main circuit board. It is held in place by a metal clip. It prevents water and other liquids from contaminating the flow system. Replace the filter when it appears dirty, discolored, or clogged. If a liquid other than water is drawn into the filter, replace the filter as soon as possible.

### ***Charcoal Filter***

The charcoal filter is located above the LEL transmitter. It is held in place by a metal clip. The charcoal filter is placed after the H<sub>2</sub>S sensor and before the CO sensor in the flow system. It scrubs out interfering gasses which may cause the CO sensor to respond, such as H<sub>2</sub>S or certain hydrocarbons. Replace the charcoal filter when false high CO readings are noticed, especially in the presence of H<sub>2</sub>S.

### ***Pump***

The pump is located to the left of the main circuit board near the bottom left of the sample-draw detector. The pump pulls the gas sample into the sample-draw detector. The pump operates on 24 VAC, which is generated from the 24 VDC supplied to the sample draw detector.

### ***Flowmeter***

The flowmeter is attached to the main circuit board near the top left corner (see Figure 4.) A ball in the flowmeter column indicates the flow rate to the sensors. The flowmeter measures the flow in the range 0.2 to 2.0 SCFH (Standard Cubic Feet per Hour). Although the sample-draw detector will operate down to a flow of 0.6 SCFH, the optimum flow rate is 1.2 SCFH.

### ***Bypass valve***

The bypass valve is to the left of the flowmeter. The bypass valve adjusts the flow rate to the sensors. Use a flat-blade screwdriver to adjust the bypass valve.

---

**NOTE:** The bypass valve allows fine adjustments of the flow rate. For a wider range of adjustment, use the flow adjust potentiometer (see Figure 4.)

---

### ***Status lights***

Two status lights are above the flowmeter.

#### ***Pilot light***

The green Pilot light is on when the sample-draw detector is receiving power.

#### ***Fail light***

The red Fail light is on when the sample flow rate is below the low flow level.

---

**NOTE:** The factory set low flow level is 0.6 SCFH ( $\pm 0.2$ ). See “Adjusting the Low Flow Setting” on page 19 to adjust this setting.

---

### ***Pressure switch***

The pressure switch is mounted to the opposite side of the main circuit board. The pressure switch monitors the flow rate to the sensors.

If the flow rate falls below the preset low flow level, the pressure switch causes the fail relay to interrupt the signal in the 4-20 mA line for the CO channel. This causes a downscale reading at the monitor on this channel. The low flow level is factory-set at 0.6 SCFH ( $\pm 0.2$  SCFH).

---

**NOTE:** There is no low flow indication for the LEL and oxygen channels.

---

### ***Flow Block***

The flow block is located in the lower right corner of the sample-draw detector. All the sensors are installed in the flow block. The flow block routes the sampled air to each sensor.

### ***EXHAUST fitting***

The EXHAUST fitting on the bottom of the housing allows the gas sample to exit the sample-draw detector. The EXHAUST fitting accepts 1/4 in. rigid tubing. See the Installation section on page 11 to connect tubing to the EXHAUST fitting.

## **Detection System**

The detection system consists of the gas sensors, LEL and oxygen transmitters, preamp circuit board, and the main circuit board.

### ***Combustible Gas Sensor***

The combustible gas sensor is installed in the lower left of the flow block. The combustible gas sensor includes the sensing elements, flame arrestor, connector, and sensor leads.

#### *Sensing elements*

Two sensing elements are protected within the sensor assembly. Through a series of thermal and electronic reactions, these elements produce an output that is proportional to the detection range of the sample draw detector. The LEL transmitter converts the output to a 4 - 20 mA signal which can be used by a recording or monitoring device.

The porous flame arrestor allows the gas sample to enter the sensor assembly and contact the sensing element. The flame arrestor also contains sparks within the sensor.

#### *Connector*

The top of the sensor includes five pins that plug into the socket connector. This connector allows you to replace the sensor without disconnecting the wiring. The sensor leads are soldered to the connector.

#### *Sensor leads*

Four color-coded leads extend from the connector. The leads allow you to connect the combustible gas sensor to the main circuit board.

### ***Oxygen sensor***

The oxygen sensor is installed in the lower right of the flow block. The oxygen sensor includes the oxygen cell, connector, and sensor leads.

#### *Oxygen cell*

The oxygen cell is protected within the sensor assembly. Through a series of chemical and electronic reactions, the cell produces a millivolt output that is proportional to the detection range of the sample-draw detector. The oxygen transmitter converts the output to a 4 - 20 mA signal which can be used by a recording or monitoring device.

#### *Connector*

The cable that extends from the sensor terminates in a socket that plugs into a matching 7-pin male connector. The socket and connector allow you to replace the sensor without disconnecting the wiring. The sensor leads are soldered to the male connector.

### Sensor leads


Two color-coded leads extend from the connector. The leads allow you to connect the oxygen sensor to the main circuit board.

### **Carbon Monoxide Sensor**

The carbon monoxide gas sensor is installed in the upper left side of the flow block. It has 4 pins which mate with sockets in the preamp circuit board.

### **Hydrogen sulfide gas sensor**

The hydrogen sulfide gas sensor position is located in the upper right side of the flow block. In the 35-3010RKA-03, the hydrogen sulfide gas sensor position in the flow block is occupied by a dummy sensor.



**Figure 3: LEL & Oxygen Transmitters**

### **LEL Transmitter**

The LEL transmitter is mounted to the left of the oxygen transmitter and above the main circuit board. It consists of the span pot, zero pot, two internally wired terminal strips, and the test points.

### Span/zero pots

The span and zero pots are located at the bottom edge of the transmitter and are used for calibration. Use the span pot to make adjustments to gas response readings and the zero pot to make adjustments to the zero reading.

### Transmitter interconnect terminal strip

The transmitter interconnect terminal strip is the six-point terminal strip near the right edge of the transmitter. The transmitter is factory wired to the sensor and main circuit board.

#### Test points

The test points are located on the left side of the transmitter and are labeled **TP+** and **TP-**. A 100 mV - 500 mV output is available at these test points for use during calibration.

#### ***Oxygen Transmitter***

The oxygen transmitter is mounted to the right of the LEL transmitter and above the main circuit board. The amplifier includes the amplifier type selector, two internally wired terminal strips, span pot, zero pot, and test points.

#### Transmitter type selector

The transmitter type selector is near the bottom left corner of the amplifier. It is to the left of the detector terminal strip and below the span pot.

The transmitter included with the sample-draw detector is designed for use with RKI's oxygen and toxic gas sensors. The transmitter type selector determines for which sensor the amplifier is intended. For oxygen transmitters, a jumper block is installed over the **OXY** selector.

#### Detector terminal strip

The detector terminal strip is the four-point terminal strip near the bottom edge of the transmitter. It is factory wired to the transmitter.

#### Interconnect terminal strip

The interconnect terminal strip is the four-point terminal strip near the top edge of the transmitter. It is factory wired to the main circuit board.

#### Span/zero pots

The span and zero pots are located on the left side of the transmitter and are used for calibration. Use the span pot to make adjustments to gas response readings and the zero pot to make adjustments to the zero reading.

#### Test points


The test points are located on the top of the transmitter on either side of the transmitter interconnect terminal strip. They are labeled **TP+** and **TP-**. A 100 mV - 500 mV output is available at these test points for use during calibration.

#### ***Preamp Circuit Board***

The preamp circuit is used to connect the CO and H<sub>2</sub>S sensors to the main circuit board and to secure the sensors in the flow block. Two cables mate to the main circuit board: the one on the left is for the CO sensor signal and the one on the right is for the H<sub>2</sub>S sensor signal.



## Main Circuit Board



**Figure 4: Main Circuit Board**

The main circuit board includes the interconnect terminal strip, sensor/transmitter terminal strip, amp 1 circuit, amp 2 circuit, pump terminal strip, and relay.

**NOTE:** The flowmeter and status lights are mounted to the main circuit board but are considered part of the flow system.

### Interconnect terminal strip

The interconnect terminal strip is the sixteen-point terminal strip near the bottom edge of the main circuit board. Use the interconnect terminal strip to connect the sample-draw detector to power and an external device.

### Sensor/Transmitter terminal strip

The sensor/transmitter terminal strip is the sixteen-point terminal strip near the right edge of the circuit board. Use the transmitter terminal strip to connect sensors or transmitters to the main circuit board.

**NOTE:** The sensors and transmitters are factory wired to the sensor/transmitter terminal strip. See the “Installation” on page 11 for all wiring procedures related to the sample-draw detector.

#### Amp 1 and Amp 2 circuits

These circuits are located to the left of the sensor/transmitter terminal strip. They each include test points, a zero pot, and a span pot. Amp 1 is on the left and is for the CO channel. Amp 2 is on the right and is for the H<sub>2</sub>S channel.

The zero and span pots are used during calibration. Use the span pot to make adjustments to gas response readings and the zero pot to make adjustments to the zero reading

The test points are labeled **CAL-1 and CAL+1** for the CO channel and **CAL-2 and CAL+2** for the H<sub>2</sub>S channel. A 100 mV - 500 mV output is available at the CO of test points for use during calibration. No output is available at the H<sub>2</sub>S test points.

#### Pump terminal strip

The pump terminal strip is the four-point terminal in the top left corner of the circuit board. Use the pump terminal strip to connect the pump and pressure switch to the main circuit board.

---

**NOTE:** The pump and pressure switch are factory-wired to the circuit board. See “Wiring the Sample-Draw Detector” on page 12 for all wiring procedures related to the sample-draw detector.

---

#### Relay

The relay is approximately in the middle of the circuit board. The relay is a four pole, double-throw (4PDT) relay and is rated for 2 amps at 25 VDC (resistive). If the pressure switch senses a low flow condition, the relay interrupts the 4-20 mA signal from the CO channel which will cause a downscale reading at the controller or recording device.

---

**NOTE:** There is no flow fail indication for the LEL and oxygen channels on the 35-3010RKA-03 Sample Draw Adapter.

---

---


## Installation

This section describes procedures to mount the sample-draw gas detector in the monitoring environment and wire the sample-draw detector to power and an external device.

### Mounting the Sample-Draw Combustible Gas Detector

1. Select the mounting site. Consider the following when you select the mounting site.
  - Is there enough room to open the housing door and make wiring connections at the bottom of the housing and tubing connections at the right of the housing?

Make sure there is sufficient room to perform start-up, maintenance, and calibration procedures.



**Figure 5: Mounting the Sample-Draw Detector**

2. Close and latch the housing door.

---

**NOTE:** The sample-draw detector is shipped with the mounting feet “tucked under” the housing to protect the mounting feet during shipment.

---

3. Slightly loosen the screw that secures one of the mounting feet to the housing, then rotate the mounting foot 180 degrees.
4. Tighten the screw that secures the mounting foot to the housing.
5. Repeat steps 3 and 4 for the remaining three mounting feet.

6. Position the sample-draw housing on a vertical surface at eye level (4 1/2 to 5 feet from the floor).
7. Insert 1/4 in. or 5/16 screws through the slots in the mounting feet to secure the housing to the mounting surface.

### Connecting the Sample Lines to the Sample-Draw Detector

1. Attach 1/4 in. O.D. rigid polypropylene or rigid Teflon sample tubing to the INLET fitting.

---

**CAUTION:** *If you use **flexible** sample tubing (polyurethane is acceptable), use an appropriate insert to seal the connection between the tubing and the INLET fitting.*

---

2. Place the opposite end of the tubing at the sampling area.

---

**CAUTION:** *Avoid loops or slumps in the incoming sample line. To reduce response time, keep the incoming sample line as short as possible.*

---

3. Attach rigid sample tubing to the EXHAUST fitting.
4. Route the opposite end of the tubing to an open area where the sample can safely disperse.

### Wiring the Sample-Draw Detector

---

**WARNING:** *Always verify that the power source is OFF before you make wiring connections.*

---


1. Unlatch and open the housing door of the sample-draw detector.
2. Guide a eight-conductor 18 gauge, shielded cable or eight 18 gauge wires in conduit through one of the conduit hubs at the bottom of the sample-draw housing. If necessary, use both hubs to bring the wires in making sure that all the wires for a particular channel go through the same hub.
3. Connect the cable to the sample-draw detector's interconnect terminal strip as shown in Figure 6.
4. Close and latch the housing door of the sample-draw detector.

---


**CAUTION:** *Leave the cable shield drain wire insulated and disconnected at the sample-draw detector. You will connect the opposite end of the drain wire at the device.*

---

5. Route the cable or wires in conduit leading from the sample-draw detector to the monitoring device and power.
6. Connect the drain wire to an available chassis ground at the device end.



**Figure 6: External (field) Wiring, Sample-Draw Detector**



**Figure 7: Internal (factory) Wiring, Sample-Draw Detector**

---

## Start Up

This section describes procedures to start up the sample-draw detector and place the sample-draw detector into normal operation.

### Introducing Incoming Power

1. Complete the installation procedures described earlier in this manual.
2. Verify that the power/device wiring is correct and secure.
3. Turn on or plug in the incoming power.
4. Verify that the Pilot light is on.
5. Verify that the flowmeter indicates a flow rate of approximately 1.2 SCFH. If necessary, use the bypass valve or flow adjust potentiometer to adjust the flow rate.

---

**NOTE:** The following step tests for leaks in the sample line. This test may cause a low flow condition at the sample-draw detector.

---

6. Verify that the incoming sample line is not leaking. To test the sample line, plug the open end of the sample line with your thumb. If the flowmeter ball drops to the bottom of the flowmeter, the incoming sample line is not leaking.
7. Remove your thumb from the sample line and verify the flowmeter returns to a normal flow rate.

### Setting the Zero Reading

---

**CAUTION:** *If you suspect the presence of combustible gas, carbon monoxide, or an abnormal oxygen condition (not 20.9%) in the monitoring environment, use the calibration kit and the zero air calibration cylinder to introduce “fresh air” to the sample draw adapter and verify an accurate zero setting.*

---

1. Verify that the sample-draw detector is sampling a fresh air environment (environment known to be free of combustible gas, CO and of normal oxygen content, 20.9%).
2. Open the housing door.
3. Set a voltmeter to measure in the millivolt (mV) range.
4. Check the zero reading for each channel.

For the LEL channel, plug the voltmeter leads into the test points on the LEL transmitter. Plug the positive lead into the test point labeled **TP+**; plug the negative lead into the test point labeled **TP-**.

For the oxygen channel, plug the voltmeter into the test points on the oxygen transmitter. Plug the positive lead into the test point labeled **TP+**; plug the negative lead into the test point labeled **TP-**.

For the CO channel, plug the voltmeter into the test points in the AMP 1 section of the main circuit board. Plug the positive lead into the test point labeled **CAL+1**; plug the negative lead into the test point labeled **CAL-1**.

5. Verify a voltmeter reading of 100 mV ( $\pm 2$  mV).
6. If necessary, use a small flat-blade screwdriver to adjust the zero pot until the voltmeter reading is 100 mV ( $\pm 2$  mV).
7. Close the housing door.

---

## Maintenance

This section describes maintenance procedures. It includes preventive maintenance procedures. This section also includes procedures to troubleshoot the sample-draw detector, replace components of the sample-draw detector, and adjust the low flow setting.

### Preventive Maintenance

This section describes a preventive maintenance schedule to ensure the optimum performance of the sample-draw detector. It includes daily, monthly, and quarterly procedures.

#### *Daily*

1. Verify that the pilot light is on.
2. Verify that the flowmeter indicates a flow rate of approximately 1.2 SCFH. If necessary use the bypass valve or flow adjust potentiometer to adjust the flow rate to 1.2 SCFH.
3. Verify a reading of 0%LEL for the combustible channel (100 mV at the LEL transmitter test points), 20.9% for the oxygen channel (434 mV at the oxygen transmitter test points), and 0 ppm for the CO channel(100 mV at the amp 1 test points). Investigate significant changes in the reading.

#### *Monthly*

This procedure describes a test to verify that the sample-draw detector responds properly to the target gases.

#### Preparing for the response test

---

**CAUTION:** *This procedure may cause alarms at the monitoring device. Take appropriate action to avoid this, such as entering the calibration mode at the monitoring device or disabling external alarms.*

---

1. Verify that the monitoring device is reading 0 for the combustible and CO channels and 20.9 for the oxygen channel.  
  
If the reading is not 0 on the combustible or CO channels or 20.9 on the oxygen channel, set the zero reading as described in the “Start Up” section on “Start Up” on page 15, then continue this procedure.
2. Assemble the calibration kit as described in “Assembling the Calibration Kit” on page 20. Use of a 3-gas cylinder is recommended so that all channels may be checked at once.

#### Performing the response test

---

**NOTE:** This procedure describes the RKI calibration kit that includes a gas collection bag. A calibration kit that uses a demand flow regulator is also available.

---

1. Fill the gas bag with calibration gas.
2. Connect the calibration tubing from the gas collection bag to the inlet line at or near the INLET fitting.  
  
The sample-draw detector’s pump automatically begins pulling the test sample from the gas collection bag when you connect the tubing to the inlet line.
3. After approximately one minute, verify that the reading for each channel at the monitoring



device stabilizes within  $\pm 10\%$  of the concentration of the test sample. If the reading is not within  $\pm 10\%$  of the test sample, calibrate the sample-draw detector as described in the Calibration section of this manual.

4. Remove the calibration tubing from the inlet line, then reconnect the inlet line.
5. Store the calibration kit in a safe place.

### ***Quarterly***

Calibrate the sample-draw detector as described in “Calibration” on page 20.

## **Troubleshooting**

The troubleshooting guide describes symptoms, probable causes, and recommended action for problems you may encounter with the sample-draw hydrogen sulfide gas detector.

---

**NOTE:** This troubleshooting guide describes sample-draw detector problems only. See the instruction manual for the monitoring device if it exhibits any problems.

---

### ***Fail condition***

#### ***Symptoms***

- The sample-draw detector’s Fail light is on.
- The monitoring device is operating properly but indicates a reading well below zero on one or more channels.

#### ***Probable causes***

- The sample-draw detector’s flow rate is too low because of an obstructed sample line, failed pump, etc.
- The sample-draw detector is malfunctioning.
- The sensor or transmitter wiring is disconnected or misconnected.

#### ***Recommended action***

1. At the sample-draw detector, set the correct flow rate with the bypass valve or flow adjust potentiometer.
2. If you cannot set the correct flow rate, check the sample lines for obstructions or kinks.
3. Verify that the sensor and transmitter wiring are correct and secure. The Installation section on page 11 describes detector wiring connections.
4. Calibrate the problem channel or channels as described in the Calibration section on page 20.
5. If the fail condition continues, replace the sensor from the problem channel or channels as described later in this section.
6. If the fail condition continues, contact RKI Instruments, Inc., for further instruction.

### ***Slow or no response/difficult or unable to calibrate***

#### ***Symptoms***

- One or more of the sensors respond slowly or does not respond during the monthly response test.
- Unable to accurately set the zero or response reading on one or more on the channels during the calibration procedure.
- One or more of the sensors requires frequent calibration.

---

**NOTE:** Under “normal” circumstances, the sample-draw detector requires calibration once a quarter. Some applications may require a more frequent calibration schedule.

---

*Probable causes*

- The calibration cylinder is low, out-dated, or defective.
- The sample-draw detector’s flow rate is too low because of an obstructed sample line, failed pump, etc.
- The sample-draw detector is malfunctioning.

*Recommended action*

1. Verify that the calibration cylinder contains an adequate supply of a fresh test sample.
2. If necessary, set the correct flow rate with the bypass valve or flow adjust potentiometer.
3. If you cannot set the correct flow rate, check the sample line for obstructions or kinks.
4. If the calibration/response difficulties continue, replace the sensor as described later in this section.
5. If the calibration/response difficulties continue, contact RKI Instruments, Inc., for further instruction.

## **Replacing Components of the Sample-draw Detector**

This section includes procedures to replace the sensor, filter, and ferrules.

### ***Replacing the combustible sensor***

1. Turn off incoming power.
2. Open the housing door of the sample-draw detector.
3. Unscrew and remove the two screws that secure the retraining plate, then lift the plate, connector, and sensor out of the housing.
4. Unplug the connector from the sensor.
5. Verify that you are using the correct replacement sensor (NC-6240 is printed on the sensor), then plug the sensor into the connector.
6. Place the sensor in the combustible gas sensor cavity, then position the retaining plate on the two standoffs.
7. Secure the retaining plate to the standoffs with the two screws you removed in step 3.
8. Turn on incoming power.
9. Calibrate the replacement sensor as described in the “Calibration” section on page 20.

### ***Replacing the oxygen sensor***

1. Turn off incoming power.
2. Open the housing door of the sample-draw detector.
3. Unscrew and remove the two screws that secure the retraining plate, then lift the plate, connector, and sensor out of the housing.
4. Unplug the connector from the socket that leads from the sensor.
5. Plug the socket of the replacement sensor into the connector.
6. Place the sensor in the oxygen sensor cavity, then position the retaining plate on the two standoffs.
7. Secure the retaining plate to the standoffs with the two screws you removed in step 3.

8. Turn on incoming power.
9. Calibrate the replacement sensor as described in the “Calibration” section on page 20.

#### ***Replacing the carbon monoxide sensor***

1. Turn off incoming power.
2. Open the housing door of the sample-draw detector.
3. Unscrew the 5 screws that retain the preamp circuit board.
4. Lift the preamp circuit board away from the flow block.

Be careful not to pull on the cables that connect the preamp circuit to the to the main circuit board.

There is a foam gasket in the bottom of each flow cavity beneath the circuit board. Make sure the gaskets stay in place.

5. Pull the CO sensor off the preamp circuit board. It is located in the amp 1 position (left side) of the preamp circuit board.
6. Plug the new sensor into the preamp board.
7. Reinstall the preamp circuit board with the sensors onto the flow block.
8. Turn on incoming power.
9. Calibrate the replacement sensor as described in the “Calibration” section on page 20.

#### **Adjusting the Low Flow Setting**

The factory-set low flow setting is 0.6 SCFH ( $\pm 0.2$ ). To adjust the low flow setting:

1. Use the flow adjust potentiometer (VR1) to set the flow to 0.6 SCFH.  
If the sample-draw detector goes into low flow alarm before you can adjust the flow down to 0.6 SCFH, adjust the low flow potentiometer 1/4 turn clockwise, then attempt to set the flow again. Repeat this step until you are able to adjust the flow to 0.6 SCFH.
2. Slowly turn the low flow potentiometer counterclockwise just until the sample-draw detector goes into low flow alarm.

---

**NOTE:** The low flow potentiometer is accessible through a circular cutout in the main circuit board. The cutout is labeled PS1.

---

3. Verify that the low flow alarm is 0.6 SCFH ( $\pm 0.2$ ). Repeat steps 3 and 4 if necessary.
4. Use the flow adjust potentiometer (VR1) to set the flow to 1.2 SCFH.
5. Make sure the sample-draw detector’s Fail light is off.

---

## Calibration

This section describes how to calibrate the sample-draw adapter. It includes procedures to assemble the calibration kit, set the zero reading for each channel, set the response reading for each channel, and return to normal operation.

---

**NOTE:** This procedure describes calibration using a gas collection bag and a 3-gas calibration cylinder. A demand-flow regulator calibration kit is also available for calibrating the sample-draw detector.

---

### Preparing for Calibration

---

**CAUTION:** *This procedure may cause alarms at the monitoring device. Take appropriate action to avoid this, such as entering the calibration mode at the monitoring device.*

---

1. Open the housing door.
2. Set a voltmeter to measure in the millivolt (mV) range.
3. When checking the mV output of each channel, plug the voltmeter leads into the appropriate test points.

For the LEL channel, plug the voltmeter leads into the test points on the LEL transmitter. Plug the positive lead into the test point labeled **TP+**; plug the negative lead into the test point labeled **TP-**.

For the oxygen channel, plug the voltmeter into the test points on the oxygen transmitter. Plug the positive lead into the test point labeled **TP+**; plug the negative lead into the test point labeled **TP-**.

For the CO channel, plug the voltmeter into the test points in the AMP 1 section of the main circuit board. Plug the positive lead into the test point labeled **CAL+1**; plug the negative lead into the test point labeled **CAL-1**.

4. Use the following formula to determine the correct test points output for the calibrating sample.

$$\text{Output (mV)} = (\text{calibrating sample/fullscale}) \times 400 + 100$$

For example, with a calibrating sample of 50 %LEL methane and a fullscale setting of 100%LEL, the correct output for the LEL test points is 300 mV.

$$300 \text{ (mV)} = (50/100) \times 400 + 100$$

### Assembling the Calibration Kit

---

**NOTE:** If you can verify a fresh air environment, it is not necessary to use a zero air calibration cylinder to set the zero reading. Perform Step 1 and then proceed to the next section, “Setting the Zero Reading”.

---

1. Connect the calibration kit sample tubing to the fitting on the gas collection bag.
2. Connect the sample tubing from the gas collection bag to the inlet line at or near the INLET fitting.

Allow the sample-draw pump to draw out any residual gas in the gas collection bag.

3. Disconnect the calibration kit sample tubing from the inlet line.

4. Close the clamp right away. The clamp is attached to the calibration kit sample tubing.
5. Connect the tubing from the gas collection bag to the fixed flow regulator, then open the clamp.
6. Screw the fixed flow regulator onto the zero air calibration cylinder.
7. Turn the on/off knob on the regulator counterclockwise to open it. The gas collection bag begins to fill.
8. When the bag is full, turn the on/off knob on the regulator clockwise to close it.
9. Close the clamp, then disconnect the sample tubing from the fixed flow regulator.
10. Unscrew the fixed flow regulator from the cylinder.

### Setting the Zero Reading

1. Open the clamp, then connect the sample tubing from the gas collection bag to the sample-draw detector's inlet line. **This step is not necessary if you verified a fresh air environment earlier in this procedure.**
2. Allow the reading to stabilize for approximately 1 minute.
3. Verify a voltmeter reading of  $100\text{ mV} \pm 2\text{ mV}$  (or 434 mV for oxygen) at the test points for each channel as described in the Preparing for Calibration section above.
4. If necessary, use a small flat-blade screwdriver to adjust the zero pot for the appropriate channel until the voltmeter reading is  $100\text{ mV} \pm 2\text{ mV}$  (or 434 mV for oxygen).
5. If you did not use zero air to set the zero reading, connect the sample tubing from the sample bag to the sample-draw detector's inlet line at or near the sample-draw detector's INLET fitting.
6. Allow the sample-draw pump to draw out any residual gas in the gas collection bag.
7. Disconnect the sample tubing from the inlet line, then close the clamp.
8. Connect the sample tubing from the gas collection bag to the fixed flow regulator, then open the clamp.
9. Screw the fixed flow regulator onto the calibration gas cylinder.
10. Turn the on/off knob on the regulator counterclockwise to open it. The gas collection bag begins to fill.
11. When the bag is full, turn the on/off knob on the regulator clockwise to close it.
12. Close the clamp, then disconnect the sample tubing from the fixed flow regulator.
13. Unscrew the fixed flow regulator from the cylinder.

### Setting the Response Reading

1. Open the clamp, then connect the sample tubing from the gas collection bag to the inlet line at or near the sample-draw detector's INLET fitting.
2. Allow the sample-draw detector to respond to the calibrating sample for 1 minute.
3. After one minute, check the mV output on the LEL transmitter test points and verify that the reading matches the response reading ( $\pm 2\text{ mV}$ ) you determined earlier.
4. If necessary, use the span pot on the LEL transmitter to adjust the reading to match the correct response reading.
5. Repeat steps 3 and 4 for the oxygen channel and the CO channel using the appropriate test points described in the Preparing For Calibration section above and the appropriate span pot.
6. Allow the sample-draw pump to draw out any residual gas in the gas collection bag.

7. Disconnect the sample tubing from the sample-draw detector's inlet line, then close the clamp.
8. Reconnect the incoming sample line.
9. Wait 1 to 2 minutes to allow the calibration gas reading to be drawn out and the readings to stabilize.
10. Store the components of the calibration kit in a safe and convenient place.

---

## Parts List

Table 2 lists replacement parts and accessories for the sample-draw gas detector.

**Table 2: Parts List**

| Part Number  | Description  |
|--------------|--|
| 06-1248RK    | Sample tubing, 3/16 x 5/16, specify length, (for calibration kit)                |
| 07-0034RK    | Sealing gasket, for CO and H <sub>2</sub> S flow block cavities                  |
| 30-0610RK    | Pump   |
| 33-0171RK    | Hydrophobic filter (AcroPak)   |
| 33-6095RK    | Charcoal filter, CF-188  |
| 61-0145RK    | Combustible sensor   |
| 65-0601RK    | Oxygen sensor  |
| 81-0090RK-03 | 3-gas calibration gas cylinder, 50% LEL methane/12% oxygen/50 ppm CO, 103 liters |
| 81-0076RK-03 | Zero air calibration cylinder (103 liter)  |
| 81-1051RK-60 | Regulator, w/gauge and knob, fixed flow, 6 LPM                                   |
| 81-1126RK    | Gas collection bag (2 liter)   |
| ES-1531-CO   | CO sensor  |